

PROVISIONAL CERTIFICATE OF ANALYSIS

13X 32180 (batch A)

Reference Material Information

Type: AUSTENITIC STAINLESS STEEL (WROUGHT)

Form and Size: Disc, 40mm diameter

Produced by: Instytut Metalurgii Zelaza, Poland

Certified and supplied by: MBH Analytical Ltd

Provisional Composition

Percentage element by weight

Element	C	Si	S	P	Mn	Ni	Cr	Mo
%	0.03	0.50	0.01	0.005	2.1	10.0	18.8	0.24

Element	Cu	Co	V	W	Al	Ti	Sn	N
%	0.50	0.035	0.025	0.035	0.025	0.81	0.01	0.01

Notes

The above values are based on the very early results from an incomplete interlaboratory testing programme which will be fully detailed on the final certificate. The stated values should be used for guidance only. This material should not be used for calibration purposes until formal certification is completed.

This document will be superseded in December 2020, and should not be used after that time.

This material will remain stable indefinitely, provided adequate precautions are taken to protect it from cross-contamination, extremes of temperature and atmospheric moisture.

The material to which this provisional certificate of analysis refers is supplied subject to our general conditions of sale.

Certified by:

MBH ANALYTICAL LIMITED _____

on 20th September 2019