

# PROVISIONAL CERTIFICATE OF ANALYSIS

**31X 7835.3 (batch K)**

## Reference Material Information

Type: LEADED BRASS (CHILL CAST)  
Form and Size: Disc ~40mm diameter  
Produced by: Maybrey Reliance Foundry  
Certified and Supplied by: MBH Analytical Ltd

## Provisional Composition

### Percentage element by weight

Element	Sn	Pb	Zn	Fe	Ni	Al	Si	As
%	0.36	1.70	36.6	0.49	0.15	0.48	0.075	0.058

Element	Sb	P	Bi	Co	Ag	Te	Cd	Cu
%	0.061	0.035	0.029	0.006	0.021	0.001	0.006	60.0

## Notes

The above values are based on the early results from an incomplete interlaboratory testing programme, which will be fully detailed on the final certificate. The stated values should be used for guidance only. This material should not be used for calibration purposes until formal certification is completed.

This document will be superseded in June 2019, and should not be used after that time.

This material will remain stable indefinitely, provided adequate precautions are taken to protect it from cross-contamination, extremes of temperature and atmospheric moisture.

The material to which this provisional certificate of analysis refers is supplied subject to our general conditions of sale.

## Certified by:

MBH ANALYTICAL LIMITED \_\_\_\_\_

on 17<sup>th</sup> January 2019

C Eveleigh