

# PROVISIONAL CERTIFICATE OF ANALYSIS

**31X B5 (batch L)**

## Reference Material Information

Type: TRACES IN BRASS (CHILL CAST)

Form and Size: Disc ~40mm diameter

Produced by: Polycast Ltd

Certified and Supplied by: MBH Analytical Ltd

## Provisional Composition

### Percentage element by weight

Element	Sn	Pb	Zn	Fe	Mn	Ni	Si	As
%	0.27	0.085	24.0	0.035	0.002	0.027	(0.001)	0.035

Element	Cr	Co	Al	Bi	Sb	Cd	B	Cu
%	(0.001)	0.025	0.015	0.008	0.015	0.004	(0.0005)	(rem)

## Notes

The above values are based on the results from the very early stages of an incomplete interlaboratory testing programme, which will be fully detailed on the final certificate. The stated values should be used for guidance only. This material should not be used for calibration purposes until formal certification is completed.

This document will be superseded in June 2018 and should not be used after that time.

This material will remain stable indefinitely, provided adequate precautions are taken to protect it from cross-contamination, extremes of temperature and atmospheric moisture.

The material to which this provisional certificate of analysis refers is supplied subject to our general conditions of sale.

## Certified by:

MBH ANALYTICAL LIMITED \_\_\_\_\_

on 30<sup>th</sup> June 2017

C Eveleigh

