

PROVISIONAL CERTIFICATE OF ANALYSIS

32X LB10 (batch G)

Reference Material Information

Type: LEADED BRONZE (CHILL CAST)
Form and Size: Disc ~40mm diameter
Produced by: Polycast Ltd
Certified and supplied by: MBH Analytical Ltd

Provisional Composition

Percentage element by weight

Element	Sn	Pb	Zn	Fe	Ni	Al	Si	As
%	8.25	12.5	0.110	<0.005	0.69	<0.001	<0.001	0.165

Element	Co	Bi	Sb	Ag	P	S	Te	Cu
%	0.084	0.092	0.60	0.07	0.003	0.01	0.010	77.4

Notes

The above values are derived from the early results of an incomplete interlaboratory testing programme, which will be fully detailed on the final certificate. The stated values should be used for guidance only. This material should not be used for calibration purposes until formal certification is completed.

This document will be superseded in February 2019, and should not be used after that time.

This material will remain stable indefinitely, provided adequate precautions are taken to protect it from cross-contamination, extremes of temperature and atmospheric moisture.

The material to which this provisional certificate of analysis refers is supplied subject to our general conditions of sale.

Certified by:

MBH ANALYTICAL LIMITED _____

on 21st September 2018

C Eveleigh