

PROVISIONAL CERTIFICATE OF ANALYSIS

36X CN2 (batch K)

Reference Material Information

Type: CUPRO-NICKEL (CHILL CAST)
Form : Chips, or disc ~40mm diameter
Produced by: Polycast Ltd
Certified and supplied by: MBH Analytical Ltd

Provisional Composition

Percentage element by weight

Element	Sn	Pb	Zn	Fe	Ni	Co	Si	Cr
%	0.024	0.05	1.05	0.045	11.2	0.21	0.05	0.004

Element	Mn	Ti	P	S	C	Mg	Cu
%	0.75	0.04	0.045	0.009	0.001	0.035	(bal)

Notes

These provisional values are derived from the very early stages of an incomplete interlaboratory testing programme, which will be fully detailed on the final certificate. The stated values should be used for guidance only. This material should not be used for calibration purposes until formal certification is completed.

This document will be superseded in June 2018, and should not be used after that time.

This material will remain stable indefinitely, provided adequate precautions are taken to protect it from cross-contamination, extremes of temperature and atmospheric moisture.

The material to which this provisional certificate of analysis refers is supplied subject to our general conditions of sale.

Certified by:

MBH ANALYTICAL LIMITED _____

on 29th September 20157

C Eveleigh