

# PROVISIONAL CERTIFICATE OF ANALYSIS

**42X Z7 (batch D)**

## Reference Material Information

Type: ZINC / ALUMINIUM (CAST)

Form and Size: Disc ~50mm diameter

Manufactured by: MBH Analytical Ltd

Certified and Supplied by: MBH Analytical Ltd

## Provisional Composition

### Percentage element by weight

<b>Element</b>	<b>Pb</b>	<b>Mg</b>	<b>Al</b>	<b>Cd</b>	<b>Fe</b>	<b>Cu</b>
%	0.004	0.010	4.35	0.002	0.025	0.025

<b>Element</b>	<b>Sn</b>	<b>Ni</b>	<b>Mn</b>	<b>Si</b>	<b>Ce</b>	<b>La</b>
%	0.0025	0.008	0.005	0.002	0.060	0.025

## Notes

These provisional values are derived from the very early results of an incomplete interlaboratory testing programme, which will be fully detailed on the final certificate. The stated values should be used for guidance only. This material should not be used for calibration purposes until formal certification is completed.

This document will be superseded in June 2019, and should not be used after that time.

This material will remain stable indefinitely, provided adequate precautions are taken to protect it from cross-contamination, extremes of temperature and atmospheric moisture.

The material to which this provisional certificate of analysis refers is supplied subject to our general conditions of sale.

## Certified by:

MBH ANALYTICAL LIMITED \_\_\_\_\_

on 1st November 2018