

# PROVISIONAL CERTIFICATE OF ANALYSIS

**51X G00H4 (batch C)**

## Reference Material Information

Type: ALUMINIUM WITH IMPURITIES (CAST)  
Form and Size: Disc ~50mm diameter  
Manufactured by: MBH Analytical Ltd  
Certified and Supplied by: MBH Analytical Ltd

## Provisional Composition

### Percentage element by weight

Element	Cu	Mg	Si	Fe	Mn	Ni	Zn	Pb	Sn	Ag
Value	0.049	0.054	0.025	0.082	0.080	0.033	0.21	0.020	0.029	0.021

Element	Ti	Cr	Co	V	Zr	Bi	Sb	Cd	Ga	Be
Value	0.030	0.061	0.011	0.021	0.028	0.025	0.008	0.027	0.033	0.0012

## Notes

These provisional values are derived from the early stages of an incomplete interlaboratory testing programme, which will be fully detailed on the final certificate. The stated values should be used for guidance only. This material should not be used for calibration purposes until formal certification is completed.

This document will be superseded in August 2018, and should not be used after that time.

This material will remain stable provided adequate precautions are taken to protect it from cross-contamination, extremes of temperature and atmospheric moisture.

The material to which this provisional certificate of analysis refers is supplied subject to our general conditions of sale.

## Certified by:

MBH ANALYTICAL LIMITED \_\_\_\_\_

on 2<sup>nd</sup> May 2018

C Eveleigh