

# PROVISIONAL CERTIFICATE OF ANALYSIS

## 54X G231H1 (batch F)

### Reference Material Information

Type: ALUMINIUM / SILICON (HIPped)  
Form and Size: Disc ~65mm diameter  
Manufactured by: RSP, Holland  
Certified and Supplied by: MBH Analytical Ltd

### Provisional Composition

#### Percentage element by weight

Element	Cu	Mg	Si	Fe	Mn	Ni	Zn
Value	1.08	0.34	9.52	0.89	0.011	0.295	0.62

Element	Pb	Sn	Ti	Cr	Ca	Sr	Be
Value	0.150	0.17	0.029	0.117	0.008	0.0042	<0.0005

### Notes

These provisional values are derived from an incomplete interlaboratory testing programme, which will be fully detailed on the final certificate. The stated values should be used for guidance only. This material should not be used for calibration purposes until formal certification is completed.

This document will be superseded in July 2018, and should not be used after that time.

This material will remain stable indefinitely, provided adequate precautions are taken to protect it from cross-contamination, extremes of temperature and atmospheric moisture.

The material to which this provisional certificate of analysis refers is supplied subject to our general conditions of sale.

### Certified by:

MBH ANALYTICAL LIMITED \_\_\_\_\_

on 15th May 2018

C Eveleigh