

PROVISIONAL CERTIFICATE OF ANALYSIS

55X A30J1 (batch J)

Reference Material Information

Type: ALUMINIUM / SILICON / COPPER
(Wrought, pressed powder)

Form and Size: Disc 65mm diameter x 25mm thick

Manufactured by: RSP Technologies, Holland

Certified and Supplied by: MBH Analytical Ltd

Provisional Composition

Percentage element by weight

Element	Cu	Mg	Si	Fe	Mn	Ni	Zn	Pb
Value	5.00	0.85	14.8	1.1	0.10	0.001	0.14	0.15

Element	Sn	Ti	Cr	Co	Bi	Be	Ca	P
Value	0.010	0.21	0.010	0.077	0.018	0.005	0.003	0.006

Notes

These provisional values are derived from the very early results of an incomplete interlaboratory testing programme, which will be fully detailed on the final certificate. The stated values should be used for guidance only. This material should not be used for calibration purposes until formal certification is completed.

This document will be superseded in June 2019, and should not be used after that time.

This material will remain stable indefinitely, provided adequate precautions are taken to protect it from cross-contamination, extremes of temperature and atmospheric moisture.

The material to which this provisional certificate of analysis refers is supplied subject to our general conditions of sale.

Certified by:

MBH ANALYTICAL LIMITED _____

on 1st November 2018

C Eveleigh