

PROVISIONAL CERTIFICATE OF ANALYSIS

55X G02D9 (batch L)

Reference Material Information

Type: ALUMINIUM / SILICON / COPPER (HIP)
Form and Size: Disc 65mm diameter
Manufactured by: RSP, Holland
Certified and Supplied by: MBH Analytical Ltd

Provisional Composition

Percentage element by weight

Element	Cu	Mg	Si	Fe	Mn	Ni	Zn	Pb
Value	3.46	0.118	8.15	0.615	0.205	0.52	2.21	0.147

Element	Sn	Ti	Cr	V	Co	Sb	Cd	Be
Value	0.28	0.105	0.088	0.017	0.0058	0.047	0.0015	0.0002

Notes

These provisional values are derived from the results of an incomplete interlaboratory testing programme, which will be fully detailed on the final certificate. The stated values should be used for guidance only. This product should not be used for calibration purposes until formal certification is completed.

This document will be superseded in June 2020 and should not be used after that time.

This product will remain stable indefinitely, provided adequate precautions are taken to protect it from cross-contamination, extremes of temperature and atmospheric moisture.

The sample to which this provisional certificate of analysis refers is supplied subject to our general conditions of sale.

Certified by:

MBH ANALYTICAL LIMITED _____
C Eveleigh

on 17th September 2019