

PROVISIONAL CERTIFICATE OF ANALYSIS

212X 4007 (batch B)

Reference Material Information

Type: NICKEL/COPPER MONEL (CHILL CAST)

Form and Size: Disc 40mm diameter

Produced by: Polycast Ltd

Certified and supplied by: MBH Analytical Ltd

Provisional Composition

Percentage element by weight

Element	C	Si	S	P	Mn	Cu	Fe	Nb
%	0.05	2.1	0.002	0.025	1.0	29.0	2.0	2.4

Element	Mg	Pb	Sn	Zn	Cd	Bi	Se	Ni
%	0.05	0.02	0.01	0.08	0.003	0.04	0.02	62.8

Notes

These provisional values are based on the very early results from an incomplete interlaboratory testing programme which will be fully detailed on the final certificate. The stated values should be used for guidance only. This material should not be used for calibration purposes until formal certification is completed.

This document will be superseded in September 2017, and should not be used after that time.

This material will remain stable provided adequate precautions are taken to protect it from cross-contamination, extremes of temperature and atmospheric moisture.

The material to which this provisional certificate of analysis refers is supplied subject to our general conditions of sale.

Certified by:

MBH ANALYTICAL LIMITED _____

C Eveleigh

On 23rd November 2016