

PROVISIONAL CERTIFICATE OF ANALYSIS**33X RB2 (batch B)****Certified Reference Material Information**

Type: LEADED RED BRASS (CHILL CAST)
Form and Size: Disc ~42mm diameter
Produced by: Polycast Ltd
Certified and supplied by: MBH Analytical Ltd.

Provisional Composition**Percentage element by weight**

Element	Sn	Pb	Zn	Fe	Ni	Al	Si	As	Mn
%	4.6	3.0	8.9	0.5	0.3	0.01	0.002	0.04	0.01

Element	Bi	Sb	P	S	Cr	Co	Ag	Te	Cu
%	0.09	0.05	0.04	0.07	0.003	0.03	0.09	0.007	~82

Notes

The above values are based on the early results from an incomplete interlaboratory testing programme, which will be fully detailed on the final certificate. The stated values should be used for guidance only. This material should not be used for calibration purposes until formal certification is completed.

This document will be superseded in July 2017, and should not be used after that time.

This material will remain stable provided adequate precautions are taken to protect it from cross-contamination, extremes of temperature and atmospheric moisture.

The material to which this provisional certificate of analysis refers is supplied subject to our general conditions of sale.

Certified By:MBH ANALYTICAL LIMITED _____ on 28th September, 2016

L Maxim

