

PROVISIONAL CERTIFICATE OF ANALYSIS

36X 274 (batch B)

Reference Material Information

Type: COPPER / NICKEL / SILICON (WROUGHT)
Form: Disc ~40mm diameter
Produced by: Stock Bar, UNS C18000
Certified and supplied by: MBH Analytical Ltd

Provisional Composition

Percentage element by weight

Element	Ni	Si	Cr	Fe	Co	Ag
%	2.55	0.67	0.33	0.02	0.005	0.002

Element	Sn	Pb	Zn	Al	Mn	Cu
%	0.001	0.001	0.001	0.001	0.001	[bal]

Notes

These provisional values are derived from the very early stages of an incomplete interlaboratory testing programme, which will be fully detailed on the final certificate. The stated values should be used for guidance only. This material should not be used for calibration purposes until formal certification is completed.

This document will be superseded in June 2017, and should not be used after that time.

This material will remain stable provided adequate precautions are taken to protect it from cross-contamination, extremes of temperature and atmospheric moisture.

The material to which this provisional certificate of analysis refers is supplied subject to our general conditions of sale.

Certified by:

MBH ANALYTICAL LIMITED _____

L Maxim

on 23rd November 2016