

PROVISIONAL CERTIFICATE OF ANALYSIS

54X G06H4 (batch R)

Reference Material Information

Type: ALUMINIUM / SILICON (HIPped Powder)
Form and Size: Disc ~65mm diameter
Manufactured by: RSP Technology, Holland
Certified and Supplied by: MBH Analytical Ltd

Provisional Composition

Percentage element by weight

Element	Cu	Mg	Si	Fe	Mn	Ni	Zn	Pb
Value	0.13	0.12	12.1	0.27	0.51	0.12	0.13	0.045

Element	Sn	Ti	Cr	Co	V	Ca	Cd	Li
Value	0.005	0.12	0.015	0.12	0.055	0.002	0.003	0.002

Notes

These provisional values are derived from the very early results of an incomplete interlaboratory testing programme, which will be fully detailed on the final certificate. The stated values should be used for guidance only. This material should not be used for calibration purposes until formal certification is completed.

This document will be superseded in July 2017, and should not be used after that time.

This material will remain stable provided adequate precautions are taken to protect it from cross-contamination, extremes of temperature and atmospheric moisture.

The material to which this provisional certificate of analysis refers is supplied subject to our general conditions of sale.

Certified by:

MBH ANALYTICAL LIMITED _____

C Eveleigh

on 23rd November 2016